

Index

1. Description
 - 1.1 Principle of MACS® Separation
 - 1.2 Background and product applications
 - 1.3 Reagent and instrument requirements
2. Protocol
 - 2.1 Sample preparation
 - 2.2 Magnetic labeling
 - 2.3 Magnetic separation
 - 2.4 (Optional) Evaluation of TCR γ/δ^+ T cell purity
3. Example of a separation using the TCR γ/δ^+ T Cell Isolation Kit
4. Reference

1. Description

Components	2 mL TCRγ/δ^+ T Cell Biotin-Antibody Cocktail, human: Cocktail of biotin-conjugated monoclonal anti-human antibodies against antigens not expressed by TCRγ/δ^+ T cells. 2 mL Anti-Biotin MicroBeads: MicroBeads conjugated to a monoclonal anti-biotin antibody (isotype: mouse IgG1).
Size	For 10 ⁹ total cells, up to 100 separations.
Product format	The Biotin-Antibody Cocktail is supplied in buffer containing stabilizer and 0.05% sodium azide. Anti-Biotin MicroBeads are supplied in buffer containing 0.05% sodium azide.
Storage	Store protected from light at 2–8 °C. Do not freeze. The expiration date is indicated on the vial label.

1.1 Principle of MACS® Separation

Using the TCR γ/δ^+ T Cell Isolation Kit, human, untouched TCR γ/δ^+ cells are isolated by depletion of non-TCR γ/δ^+ cells. Non-TCR γ/δ^+ cells are indirectly magnetically labeled with a cocktail of biotin-conjugated monoclonal antibodies, as primary labeling reagent, and anti-biotin monoclonal antibodies conjugated to MicroBeads, as secondary labeling reagent. The magnetically labeled non- γ/δ T cells are depleted by retaining them on a MACS® Column in the magnetic field of a MACS Separator, while the unlabeled γ/δ T cells pass through the column.

1.2 Background and product applications

The TCR γ/δ^+ T cells represent a subset of T cells expressing a T cell receptor (TCR) variant, composed of gamma and delta chains. The γ/δ TCR is expressed by 2–10% of all T cells in human peripheral blood, whereas the majority of T cells expresses α/β TCRs. γ/δ T cells possess distinct structural and antigen-binding characteristics. They specifically bind small non-peptide antigens, derived from necrotic cells or bacteria. In contrast to α/β T cells, γ/δ T cells directly recognize the antigens—independently of antigen-presenting cells (APCs) and MHC molecules. γ/δ T cells display a range of innate effector functions including rapid secretion of chemokines and cytokines as well as target cell lysis. They also contribute to the adaptive immune system, for example, they exhibit immunoregulatory and memory function. Recently, γ/δ T cells are discussed to have professional antigen-presenting capacity, initiating adaptive immune responses.¹

The TCR γ/δ^+ T Cell Isolation Kit is an indirect magnetic labeling system for the isolation of untouched TCR γ/δ^+ T cells from human peripheral blood mononuclear cells (PBMCs). Non- γ/δ T cells, i.e. α/β T cells, NK cells, B cells, dendritic cells, granulocytes, monocytes, stem cells, and erythroid cells are indirectly labeled using a cocktail of biotin-conjugated antibodies and Anti-Biotin MicroBeads. The TCR γ/δ^+ T Cell Isolation Kit provides an efficient and fast way to isolate untouched γ/δ T cells by depletion of the magnetically labeled cells.

Example applications

- Functional studies on γ/δ T cells in which effects due to antibody-cross-linking of cell surface proteins should be avoided.
- Isolation of untouched γ/δ T cells from human peripheral blood for phenotypical and functional analysis, e.g. studies on cytokine expression and receptor signaling, antigen recognition, cytotoxicity.

1.3 Reagent and instrument requirements

- Buffer: Prepare a solution containing phosphate-buffered saline (PBS) pH 7.2, 0.5% bovine serum albumin (BSA), and 2 mM EDTA by diluting MACS BSA Stock Solution (# 130-091-376) 1:20 with autoMACS™ Rinsing Solution (# 130-091-222). Keep buffer cold (4–8 °C). Degas buffer before use, as air bubbles could block the column.

▲ **Note:** EDTA can be replaced by other supplements such as anticoagulant citrate dextrose formula-A (ACD-A) or citrate phosphate dextrose (CPD). BSA can be replaced by other proteins such as human serum albumin, human serum, or fetal calf serum. Buffers or media containing Ca²⁺ or Mg²⁺ are not recommended for use.

- **MACS® Columns and MACS Separators:** Choose the appropriate MACS Separator and MACS Columns according to the number of labeled cells and to the number of total cells.

Column	Max. number of labeled cells	Max. number of total cells	Separator
MS	10 ⁷	2×10 ⁸	MiniMACS, OctoMACS, VarioMACS, SuperMACS
LS	10 ⁸	2×10 ⁹	MidiMACS, QuadroMACS, VarioMACS, SuperMACS
XS	10 ⁹	2×10 ¹⁰	SuperMACS
autoMACS	2×10 ⁸	4×10 ⁹	autoMACS

▲ **Note:** Column adapters are required to insert certain columns into VarioMACS™ or SuperMACS™ Separators. For details see the respective MACS Separator data sheet.

- (Optional) Fluorochrome-conjugated antibodies for flow-cytometric analysis, e.g. anti-TCRγ/δ-FITC, anti-TCRγ/δ-PE, anti-TCRγ/δ-APC, Anti-TCRα/β-APC (# 130-091-237), Anti-TCRα/β-PE (# 130-091-236), CD3-PE (# 130-091-374), Anti-Biotin-FITC (# 130-090-857), Anti-Biotin-PE (# 130-090-756), and Anti-Biotin-APC (# 130-090-856).
- (Optional) PI (propidium iodide) or 7-AAD for the flow-cytometric exclusion of dead cells.
- (Optional) Pre-Separation Filters (# 130-041-407) to remove cell clumps.

2. Protocol

2.1 Sample preparation

When working with anticoagulated peripheral blood or buffy coat, peripheral blood mononuclear cells (PBMCs) should be isolated by density gradient centrifugation, e.g. using Ficoll-Paque™. For details see section General protocols in the user manuals or visit www.miltenyibiotec.com/protocols.

▲ **Note:** To remove platelets after density gradient separation, resuspend cell pellet in buffer and centrifuge at 200×g for 10–15 minutes at 20 °C. Carefully aspirate supernatant. Repeat washing step.

When working with tissues, prepare a single-cell suspension by a standard preparation method. For details see section General Protocols in the user manuals or visit www.miltenyibiotec.com/protocols.

▲ **Note:** Dead cells may bind non-specifically to MACS MicroBeads. To remove dead cells, we recommend using density gradient centrifugation or the Dead Cell Removal Kit (# 130-090-101).



2.2 Magnetic labeling

▲ Work fast, keep cells cold, use pre-cooled solutions. This will prevent capping of antibodies on the cell surface and non-specific cell labeling.

▲ Volumes for magnetic labeling given below are for up to 10⁷ total cells. When working with fewer than 10⁷ cells, use same volumes as indicated. When working with higher cell numbers, scale up all reagent volumes and total volumes, accordingly (e.g. for 2×10⁷ total cells, use twice the volume of all indicated reagent volumes and total volumes).

▲ For optimal performance it is important to obtain a single-cell suspension before magnetic separation. Pass cells through 30 μm nylon mesh (Pre-Separation Filters, # 130-041-407) to remove cell clumps which may clog the column.

1. Determine cell number.
2. Centrifuge cell suspension at 300×g for 10 minutes. Aspirate supernatant completely.
3. Resuspend cell pellet in 80 μL of buffer per 10⁷ total cells.
4. Add 20 μL of Biotin-Antibody Cocktail per 10⁷ total cells.
5. Mix well and refrigerate for 10 minutes (4–8 °C).

▲ **Note:** Working on ice may require increased incubation times. Higher temperatures and/or longer incubation times lead to non-specific cell labeling.

6. Wash cells by adding 1–2 mL of buffer per 10⁷ cells and centrifuge at 300×g for 10 minutes. Aspirate supernatant completely.
7. Add 80 μL of buffer per 10⁷ total cells.
8. Add 20 μL of Anti-Biotin MicroBeads per 10⁷ total cells.
9. Mix well and refrigerate for additional 15 minutes (4–8 °C).
10. Wash cells by adding 1–2 mL of buffer per 10⁷ cells and centrifuge at 300×g for 10 minutes. Aspirate supernatant completely.
11. Resuspend up to 10⁸ cells in 500 μL of buffer.
12. Proceed to magnetic separation (2.3).

▲ **Note:** For higher cell numbers, scale up buffer volume accordingly.



2.3 Magnetic separation

▲ Choose an appropriate MACS Column and MACS Separator according to the number of total cells and the number of TCRγ/δ⁺ cells. For details see table in section 1.3.

Magnetic separation with MS and LS Columns

1. Place column in the magnetic field of a suitable MACS Separator. For details see respective MACS Column data sheet.
2. Prepare column by rinsing with appropriate amount of buffer:
MS: 500 μL LS: 3 mL
3. Apply cell suspension onto the column.
4. Allow the cells to pass through and collect effluent as fraction with unlabeled cells, representing the enriched TCRγ/δ⁺ cell fraction.
5. Wash column with appropriate amount of buffer. Perform washing steps by adding buffer three times. Only add new buffer when the column reservoir is empty.
MS: 3×500 μL LS: 3×3 mL
Collect total effluent; this is the unlabeled cell fraction.
6. (Optional) Elute retained cells outside of the magnetic field. This fraction represents the magnetically labeled non-TCRγ/δ⁺ cells.

Magnetic separation with XS Columns

For instructions on the column assembly and the separation refer to the XS Column data sheet.

Magnetic separation with the autoMACS™ Separator

▲ Refer to the autoMACS™ User Manual for instructions on how to use the autoMACS Separator.



1. Prepare and prime autoMACS Separator.
2. Place tube containing the magnetically labeled cells in the autoMACS Separator. Choose program “Depletes”.
3. Collect negative fraction from outlet port neg1. This is the TCR γ/δ^+ cell fraction.
4. (Optional) Collect positive fraction from outlet port pos1. This is the non-TCR γ/δ^+ cell fraction.

2.4 (Optional) Evaluation of TCR γ/δ^+ T cell purity

The purity of the enriched γ/δ T cells can be evaluated by flow cytometry or fluorescence microscopy. Stain aliquots of the cell fractions with fluorochrome-conjugated antibodies against TCR γ/δ . Analyze cells by flow cytometry or fluorescence microscopy. Labeling of non- γ/δ T cells with the Biotin-Antibody Cocktail can be visualized by counterstaining with a fluorochrome-conjugated anti-biotin antibody, e.g. Anti-Biotin-PE (# 130-090-756) or Anti-Biotin-APC (# 130-090-856). Staining with fluorochrome-conjugated streptavidin is not recommended.

3. Example of a separation using the TCR γ/δ^+ T Cell Isolation Kit

Isolation of untouched TCR γ/δ^+ cells from PBMCs using the TCR γ/δ^+ T Cell Isolation Kit and an LS Column. Cells are fluorescently stained with CD3-PE (# 130-091-374) and anti-TCR γ/δ -FITC. Cell debris and dead cells were excluded from the analysis based on scatter signals and PI fluorescence.



4. Reference

1. Brandes, M. *et al.* (2005) Professional antigen-presentation function by human gammadelta T Cells. *Science* 309 (5732): 264–268.

Refer to www.miltenyibiotec.com for all data sheets and protocols. Miltenyi Biotec provides technical support worldwide. Visit www.miltenyibiotec.com/local to find your nearest Miltenyi Biotec contact.

Warnings

Reagents contain sodium azide. Under acidic conditions sodium azide yields hydrazoic acid, which is extremely toxic. Azide compounds should be diluted with running water before discarding. These precautions are recommended to avoid deposits in plumbing where explosive conditions may develop.

Legal notices

Limited product warranty

Miltenyi Biotec B.V. & Co. KG and/or its affiliate(s) warrant this product to be free from material defects in workmanship and materials and to conform substantially with Miltenyi Biotec's published specifications for the product at the time of order, under normal use and conditions in accordance with its applicable documentation, for a period beginning on the date of delivery of the product by Miltenyi Biotec or its authorized distributor and ending on the expiration date of the product's applicable shelf life stated on the product label, packaging or documentation (as applicable) or, in the absence thereof, ONE (1) YEAR from date of delivery ("Product Warranty"). Miltenyi Biotec's Product Warranty is provided subject to the warranty terms as set forth in Miltenyi Biotec's General Terms and Conditions for the Sale of Products and Services available on Miltenyi Biotec's website at www.miltenyibiotec.com, as in effect at the time of order ("Product Warranty"). Additional terms may apply. BY USE OF THIS PRODUCT, THE CUSTOMER AGREES TO BE BOUND BY THESE TERMS. THE CUSTOMER IS SOLELY RESPONSIBLE FOR DETERMINING IF A PRODUCT IS SUITABLE FOR CUSTOMER'S PARTICULAR PURPOSE AND APPLICATION METHODS.

Technical information

The technical information, data, protocols, and other statements provided by Miltenyi Biotec in this document are based on information, tests, or experience which Miltenyi Biotec believes to be reliable, but the accuracy or completeness of such information is not guaranteed. Such technical information and data are intended for persons with knowledge and technical skills sufficient to assess and apply their own informed judgment to the information. Miltenyi Biotec shall not be liable for any technical or editorial errors or omissions contained herein.

All information and specifications are subject to change without prior notice. Please contact Miltenyi Biotec Technical Support or visit www.miltenyibiotec.com for the most up-to-date information on Miltenyi Biotec products.

Licenses

This product and/or its use may be covered by one or more pending or issued patents and/or may have certain limitations. Certain uses may be excluded by separate terms and conditions. Please contact your local Miltenyi Biotec representative or visit Miltenyi Biotec's website at www.miltenyibiotec.com for more information.

The purchase of this product conveys to the customer the non-transferable right to use the purchased amount of the product in research conducted by the customer (whether the customer is an academic or for-profit entity). This product may not be further sold. Additional terms and conditions (including the terms of a Limited Use Label License) may apply.

CUSTOMER'S USE OF THIS PRODUCT MAY REQUIRE ADDITIONAL LICENSES DEPENDING ON THE SPECIFIC APPLICATION. THE CUSTOMER IS SOLELY RESPONSIBLE FOR DETERMINING FOR ITSELF WHETHER IT HAS ALL APPROPRIATE LICENSES IN PLACE. Miltenyi Biotec provides no warranty that customer's use of this product does not and will not infringe intellectual property rights owned by a third party. BY USE OF THIS PRODUCT, THE CUSTOMER AGREES TO BE BOUND BY THESE TERMS.

Trademarks

autoMACS, MACS, MidiMACS, the Miltenyi Biotec logo, MiniMACS, OctoMACS, QuadroMACS, SuperMACS, and VarioMACS are registered trademarks or trademarks of Miltenyi Biotec and/or its affiliates in various countries worldwide. All other trademarks mentioned in this publication are the property of their respective owners and are used for identification purposes only.

Ficoll-Paque is a trademark of GE Healthcare companies.

Copyright © 2020 Miltenyi Biotec and/or its affiliates. All rights reserved.